

## Flame Photometers

- Industrial version supplied with Na, K, Li, Ba & Ca filters
- Clinical version supplied with Na, K & Li filters
- Low temperature, single channel
- Electronic ignition



- Flame failure safety system
- Operates on propane, butane, natural gas or L.P.G. supplies
- Fine & coarse sensitivity controls

The Jenway Models PFP7 and PFP7/C are low temperature, single channel flame photometers, designed for the routine determinations of Sodium (Na), Potassium (K), Calcium (Ca), Barium (Ba) and Lithium (Li).

Both versions are fitted with automatic flame failure detection for user safety, making them ideal for use in clinical, industrial or educational applications.



The Model PFP7/C is specifically designed for use in clinical applications. The in-built lineariser circuitry enables readings of both Na and K, at the normal clinical serum concentrations, to be displayed directly in mmol/l. Serum samples must be diluted 200:1 or 100:1 prior to presentation to the flame photometer. Jenway are able to offer the Series 7 Diluter to enable this to be carried out efficiently and accurately.

### Technical Specifications

Model	PFP7 (Industrial)	PFP7/C (Clinical)
Ranges	0-199.9	120-160mmol/l Na (linearised) 0-10.0mmol/lK
Limits of Detection	Na ≤ 0.2ppm K ≤ 0.2ppm Li ≤ 0.25ppm Ca ≤ 15ppm Ba ≤ 30ppm	– – Li ≤ 0.25ppm Ca ≤ 15ppm Ba ≤ 30ppm
Reproducibility	≤1% Coefficient of Variation for 20 consecutive samples using 10ppm Na set to read 50	
Linearity	Better than 2% when concentration of 3ppm Na/K and 5ppm Li are set to read 100	
Specificity	Interference from Na/K and Li equal in concentration to test element will be <0.5%	
Recorder Output Services	Nominal 1.00V for a reading of 100.0 Electrical: 90-125V or 190-250V @ 50/60Hz Air: moisture and oil free 6 litre/minute @ 1kg/cm <sup>2</sup> (14psi) Fuel: Propane, Butane, Natural Gas or L.P.G.	
Size	420 x 360 x 300mm	
Weight	8kg	

### Order Codes

- 500 701** Model PFP7 supplied with Na, K, Li, Ba & Ca filters, connecting hoses and clips, compressor plug, drain trap, fuses and operating/service instructions. For use on 230V 50Hz.
- 500 731** Model PFP7 supplied with Na, K, Li, Ba & Ca filters, connecting hoses and clips, compressor plug, drain trap, fuses and operating/service instructions. For use on 110V 60Hz.
- 500 801** Model PFP7/C supplied with Na, K & Li filters, connecting hoses and clips, compressor plug, drain trap, fuses and operating/service instructions. For use on 230V 50Hz.
- 500 831** Model PFP7/C supplied with Na, K & Li filters, connecting hoses and clips, compressor plug, drain trap, fuses and operating/service instructions. For use on 110V 60Hz.

### Accessories

**Please note:** These products will only operate with the correct type of air compressor and gas regulator, as specified.

- 535 001** Air compressor (220V/50Hz)  
**535 002** Air compressor (110V/60Hz)  
**037 001** Series 7 Diluter (230V/50Hz)  
**037 002** Series 7 Diluter (110V/60Hz)  
**500 176** Water separator (small)  
**500 177** Water separator (large)  
**500 134** Dust cover  
**500 172** Minor spares kit  
**500 173** Major spares kit  
**500 178** Butane regulator  
**500 179** Propane regulator  
**500 180** Natural gas regulator  
**025 016** Cleaning solution

### Clinical Standards (500ml)

- 025 008** 1.00mmol/l Li  
**025 004** 100mmol/l Na/K  
**025 006** 140mmol/l Na / 5mmol/l K  
**025 007** 120mmol/l Na / 2mmol/l K  
**025 005** 160mmol/l Na / 8mmol/l K  
**025 027** 160mmol/l Na / 80mmol/l K

### Industrial Standards (500ml)

- 025 023** 1000ppm K  
**025 024** 1000ppm Li  
**025 025** 3000ppm Ba  
**025 021** 1000ppm Na  
**025 009** 1000ppm Ca



Thank you for reading this data sheet.

For pricing or for further information, please contact us at our UK Office, using the details below.



**UK Office**

**Keison Products,**

**P.O. Box 2124, Chelmsford, Essex, CM1 3UP, England.**

**Tel: +44 (0)330 088 0560**

**Fax: +44 (0)1245 808399**

**Email: [sales@keison.co.uk](mailto:sales@keison.co.uk)**

Please note - Product designs and specifications are subject to change without notice. The user is responsible for determining the suitability of this product.